

CALIBRATION CERTIFICATE

for

Kustom LaserCam 4 Speed Meter

Customer: Wiltshire Police

Serial Number: LE3186

Certificate Number: 8112

Date of Calibration: 17 June 2024

Next Calibration: 16 June 2025

Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CD748BAF
Oscillator Frequency - 19.999652 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
Horizontal - Pass
Eye Safety - Pass

Range ± 0.2 m

Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

Speed ± 1 MPH

Actual	25 MPH	-51 MPH	73 MPH	-39 MPH	66 MPH
Indicated	25 MPH	-51 MPH	73 MPH	-39 MPH	66 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Korcz
Technical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-009

Tel: 020 8847 4400

email: info@truvelouk.com